

INSTRUCTIONS FOR USE

NUTRIENT AGAR

Ready-to-use tubes


 Nutrient Agar: *E. faecalis* and uninoculated tube

1 - INTENDED USE

In vitro diagnostic device. General purpose medium for the cultivation, sub-culture and purification of colonies of non-fastidious microorganisms, isolated from clinical and non-clinical specimens.

2 - COMPOSITION -TYPICAL FORMULA *

Beef extract	3 g
Peptone	5 g
Agar	15 g
Purified water	1000 mL

*the formula may be adjusted and/or supplemented to meet the required performances criteria.

3 - PRINCIPLE OF THE METHOD AND EXPLANATION OF THE PROCEDURE

Nutrient Agar is a culture medium based on meat peptones, used for the cultivation of non-fastidious microorganisms. The peptones provide carbon, nitrogen and vitamins for the growth of most non-fastidious microorganisms (e.g. enterobacteria, staphylococci). The absence of sodium chloride limits the swarming of *Proteus* spp. The formulation of the Nutrient Agar complies with the recommendations of ISO 65791 and ISO 10273². Nutrient Agar was one of the first media utilised in microbiology and can still be used for the examination of water and food for preparing stock cultures, for the preliminary cultivation of a sample undergoing successive bacteriological examinations, and for the isolation of microorganisms in pure culture.

4 - PHYSICAL CHARACTERISTICS

Medium appearance	very pale yellow, limpid
Final pH at 20-25°C	7.0 ± 0.2

5 - MATERIALS PROVIDED - PACKAGING

Product	Type	REF	Pack
Nutrient Agar W0104010206, EDMA: 14.01.02.01, RDM:1514035/R	Ready-to-use tubes	551810	20 glass tubes with slanted medium, 17x125 mm, flat bottom, aluminium screw-cap. Packaging: cardboard box

6 - MATERIALS REQUIRED BUT NOT PROVIDED

Incubator, laboratory equipment as required, sterile loops, swabs, ancillary culture media and reagents for the identification of the colonies.

7 - SPECIMENS

Nutrient Agar is used for the sub-culture of microorganisms isolated on other culture media and is not used for the direct inoculation of clinical samples.

8 - TEST PROCEDURE

Allow tubes to come to room temperature.

Inoculate and streak the specimen with a loop over the surface of the medium.

Routinely, incubate at 35-37°C in aerobic conditions for 18-24 hours.

The user is responsible for choosing the appropriate incubation time, temperature and atmosphere depending on the organisms to be cultivated and the local applicable protocols.

9 - READING AND INTERPRETATION

The presence of microorganisms is indicated by the appearance of colonies of various morphology and size. The characteristics of the growths are closely related to the type or types of cultivated microorganisms.

10 - USER QUALITY CONTROL

All manufactured lots of the product are released for sale after the Quality Control has been performed to check the compliance with the specifications. However, it is responsibility of the end-user to perform Quality Control testing in accordance with the local applicable regulations, in compliance with accreditation requirements and the experience of the Laboratory. Here below are listed some test strains useful for the quality control.

CONTROL STRAINS		INCUBATION T° / T / ATM	EXPECTED RESULTS
<i>S.Typhimurium</i>	ATCC 14028	35-37°C / 18-24H / A (34-38°C acc. to ISO 11133) ³	good growth
<i>E.coli</i>	ATCC 25922	35-37°C / 18-24H / A (34-38°C acc. to ISO 11133) ³	good growth
<i>Y.enterocolitica</i>	ATCC 23715	29-31°C / 18-24H / A	good growth

A: aerobic incubation; ATCC is a trademark of American Type Culture Collection



**11 - PERFORMANCES CHARACTERISTICS**

Prior to release for sale a representative sample of all lots of ready to use tubes of Nutrient Agar and of the raw material used for the production of prepared tubes (dehydrated Nutrient Agar REF 401810) are tested for productivity by comparing the results with a previously approved Reference Batch.

Productivity is tested by a quantitative method with the target strains *E.coli* ATCC 25922, *S.Typhimurium* ATCC 14028, *Y.enterocolitica* ATCC 23715; Nutrient Agar is inoculated with decimal dilutions in saline of the colonies' suspensions and incubated at 35-37°C (*E.coli*, *S.Typhimurium*) and at 29-31°C (*Y.enterocolitica*) for 18-24 hours. The colonies are enumerated on Test Batch (TB) and Reference Batch (RB) and the productivity ratio ($Pr = CFU_{TB}/CFU_{RB}$) is calculated. If Pr is $\geq 0,7$ the results are considered acceptable and conform to the specifications.

Productivity is also tested by semi-quantitative ecometric technique with *E.faecalis* ATCC 19433 and *S.aureus* ATCC 25923. After incubation at 35-37°C for 18-24 hours the amount of growth is evaluated and recorded. All strains show a good growth, comparable with the Reference Batch.

12 - LIMITATIONS OF THE METHOD

- Even if the microbial colonies are differentiated on the basis of their morphological and chromatic characteristics, it is recommended that biochemical, immunological, molecular, or mass spectrometry testing be performed on isolates, from pure culture, for complete identification. If relevant, perform antimicrobial susceptibility testing.
- This culture medium is intended as an aid in the diagnosis of infectious diseases; the interpretation of the results must be made considering the patient's clinical history, the origin of the sample and the results of other diagnostic tests.

13 - PRECAUTIONS AND WARNINGS

- This product is a qualitative *in vitro* diagnostic, for professional use only; it is to be used by adequately trained and qualified laboratory personnel, observing approved biohazard precautions and aseptic techniques.
- This product is not classified as dangerous according to current European legislation.
- This culture medium contains raw materials of animal origin. The *ante* and *post mortem* controls of the animals and those during the production and distribution cycle of the raw materials, cannot completely guarantee that these products do not contain any transmissible pathogen. Therefore, it is recommended that the ready-to-use tubes be treated as potentially infectious, and handled observing the usual specific precautions: do not ingest, inhale, or allow to come into contact with skin, eyes, mucous membranes. Download the TSE Statement from the website www.biolifeitaliana.it, describing the measures implemented by Biolife Italiana for the risk reduction linked to infectious animal diseases.
- All laboratory specimens should be considered infectious.
- The laboratory area must be controlled to avoid contaminants such as culture medium or microbial agents.
- Each tube is for single use only.
- Ready-to-use tubes of Nutrient Agar are subject to terminal sterilization by autoclaving.
- Sterilize all biohazard waste before disposal. Dispose the unused medium and the tubes inoculated with samples or microbial strains in accordance with current local legislation.
- The Certificates of Analysis and the Safety Data Sheet are available on the website www.biolifeitaliana.it.
- The information provided in this document has been defined to the best of our knowledge and ability and represents a guideline for the proper use of the product but without obligation or liability. In all cases existing local laws, regulations and standard procedures must be observed for the examination of samples collected from human and animal organic districts, for environmental samples and for products intended for human or animal consumption. Our information does not relieve our customers from their responsibility for checking the suitability of our product for the intended purpose.

14 - STORAGE CONDITIONS AND SHELF LIFE

Upon receipt, store tubes in their original pack at 2-8°C away from direct light. If properly stored, the tubes may be used up to the expiration date. Do not use the tubes beyond this date. Before use, check the integrity of the screw cap. Do not use tubes with signs of deterioration (e.g. microbial contamination, atypical colour).

15 - REFERENCES

1. ISO 6579-1:2017 Microbiology of the food chain -- Horizontal method for the detection, enumeration and serotyping of Salmonella -- Part 1: Detection of Salmonella spp.
2. ISO 10273:2017 Microbiology of the food chain -- Horizontal method for the detection of pathogenic Yersinia enterocolitica.
3. ISO 11133:2014 Microbiology of food, animal feed and water — Preparation, production, storage and performance testing of culture media.

TABLE OF APPLICABLE SYMBOLS

REF Catalogue number	LOT Batch code	IVD In vitro Diagnostic Medical Device	Manufacturer	Do not reuse	Recyclable pack This side up
Temperature limitation	Content sufficient for <n> tests	Consult Instructions for Use	Use by	Keep away from direct light	Fragile

REVISION HISTORY

Version	Description of changes	Date
Instructions for Use (IFU) - Revision 0	First publication in compliance with IVDR 2017/746	2021/04

Note: minor typographical, grammatical, and formatting changes are not included in the revision history.

